

VERIFICATION OF CALIBRATION REPORT

APR 09 2009

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130190</u> ✓
Name	<u>TIM Tuckwood</u>	ID# <u>3446</u> Date <u>4-6-09</u> ✓
A Agency	<u>Alaska State Troopers</u>	Phone # <u>895-4800</u>
Instrument Location	<u>Delta Court</u> ✓	
B Alco S/N	<u>X301596</u> ✓	Target Value <u>.081</u> ✓ High Pressure <u>600 psi</u>
Alco Test Values	<u>.082</u> ✓ 1 st Alco	<u>.082</u> ✓ 2 nd Alco
Signature	<u>Tim Tuckwood</u> ✓	
<u>003</u> <u>5/8/09</u>		

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 22 day of May 2009.

Carolyn M. Noland (Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130190

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130190

APRIL 06, 2009

OPERATOR'S NAME:

WICKWOOD TIM

OPERATOR'S NUMBER: 3446

SUBJECT'S LAST NAME:

VERIF

SUBJECT'S FIRST NAME/MI :

OF CAL

O.L. #: 12345

DEPT/AGENCY: DJT0

CASE/REPORT: 09876

TEST TYPE: U

ALCO TARGET VALUE: .001

ALCO S/N: X301596

BREATH ANALYSIS

.001 ADJUSTED FOR 28.89 in		
ALCO TARGET	.078	15:40
BLANK TEST	.000	15:41
INTERNAL STANDARD	VERIFIED	15:41
ALCO TV 28.89 in	.002	15:41
BLANK TEST	.000	15:42
SUBJECT SAMPLE	.000	15:42
BLANK TEST	.000	15:43
ALCO TV 28.90 in	.002	15:43
BLANK TEST	.000	15:44

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130190

APRIL 06, 2009

TIME 15:37

DIAGNOSTIC CHECK

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

WEIGHTS

SAMPLE CHAMBER: 50c

BREATH TUBE: 42c

BAROMETER: 28.87 in

FLOW DETECTOR: OKAY

ALCO

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#2'()*+,-./0123456789:;<=>?@BCDEFG
HIJKLMNPOQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~