

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		Datamaster cdm S/N <u>130190</u>	
Name <u>D.R. Sturkie</u>		ID# <u>3559</u>	Date <u>05-10-08</u>
Agency <u>APD</u>		Phone # <u>786-</u>	
Instrument Location <u>Anchorage JAIL</u> ✓			
Alco S/N <u>X301605</u> ✓		Target Value <u>0.080</u> ^{0.081 ccB} ✓	High Pressure <u>700</u>
Alco Test Value Average		<u>0.082</u> ✓	<u>0.081</u> ✓
		1 st Alco	2 nd Alco
Signature <u>D.R. Sturkie</u>		<u>CMB</u> <u>6/13/08</u>	

(OVER)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Daniel Sturkie, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed.



Jeanne Swartz
Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

6/16/08
Date

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130190

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130190

MAY 10, 2008 ✓ *CPB 6/13/08*

OPERATOR'S NAME:

STURKIE

OPERATOR'S NUMBER: 3559

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/NI:

L

D.L. #: L

DEPT/AGENCY: ANCI

CASE/REPORT: L

TEST TYPE: U

ALCO TARGET VALUE: .081 ✓

ALCO S/N: X301605

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 29.88 in
ALCO TARGET .080 04:31
BLANK TEST .000 04:32
INTERNAL STANDARD VERIFIED 04:32
ALCO TV 29.88 in .082 ✓ 04:33
BLANK TEST .000 04:34
SUBJECT SAMPLE .000 *CPB 6/13/08* 04:35
BLANK TEST .000 04:35
ALCO TV 29.88 in .081 04:36
BLANK TEST .000 04:36

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130190

MAY 10, 2008 ✓
TIME 04:28

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c

BAROMETER: 29.88 in

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; (=) * @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ #