VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N ~~292206W~~ 130190 PEM

Supervisor/Operator Performing the Verification Procedure:

A

Name Patrick E. Martin ID# 4072 Date 5-30-08

Agency APD Phone # 764-9592

Instrument Location Jail

B

Alco S/N 292206W Target Value .075 High Pressure 1000 +

Alco Test Value Average .070 (1st Alco) .076 (2nd Alco)

Signature PEMart CDS 6/13/08

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that,

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Patrick Martin, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

6/16/08

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

BT9 01/00 JUN 0 2 2008

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130190

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130190

MAY 30, 2008 CDB 6/13/08

OPERATOR'S NAME:
MARTIN, PATRICK E
OPERATOR'S NUMBER: 4072
SUBJECT'S LAST NAME:
MARTIN
SUBJECT'S FIRST NAME/MI :
PATRICK E
O.L. #: 1234567
DEPT/AGENCY: ANC1
CASE/REPORT: 00-00
TEST TYPE: V
ALCO TARGET VALUE: .075
ALCO S/N: 29206W

--- BREATH ANALYSIS ---

.075 ADJUSTED FOR 30.16 in

ALCO TARGET	.075	10:40
BLANK TEST	.000	10:41
INTERNAL STANDARD	VERIFIED	10:41
ALCO TV 30.16 in	.070	10:41
BLANK TEST	.000	10:42
SUBJECT SAMPLE	.000 CDB 6/13/08	10:43
BLANK TEST	.000	10:44
ALCO TV 30.16 in	.076	10:44
BLANK TEST	.000	10:45

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130190

MAY 30, 2008
TIME 10:35

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	30.16 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~