

# VERIFICATION OF CALIBRATION REPORT

SEP 18 2009

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: <u>DataMaster cdm S/N 130190</u> ✓	
Name <u>Brandi Barnett</u>	ID# <u>7038</u> Date <u>9/18/09</u> ✓
Agency <u>SCDL</u>	Phone# <u>269-5695</u>
Instrument Location <u>SCDL</u> ✓	
Alco S/N <u>SCOTTO8</u> ✓ Target Value <u>0.080</u> ✓ High Pressure <u>600psi</u>	
Alco Test Values <u>0.080</u> ✓ <u>0.080</u> ✓	
1 <sup>st</sup> Alco 2 <sup>nd</sup> Alco	
Signature <u>Brandi M. Barnett</u> ✓	
COB 9/25/09	
(OVER)	

I, Nita J. Bolz, after being first duly sworn, depose and state as follows. (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 18th day of Oct, 2009.

Carolyn M. Noland  
(Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130190

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130190

SEPTEMBER 18, 2009 ✓

OPERATOR'S NAME:

BARNETT/B

OPERATOR'S NUMBER: 7038

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

UOC

O.L. #: UOC

DEPT/AGENCY: AA01

CASE/REPORT: UOC

TEST TYPE: U

ALCO TARGET VALUE: 0.000 ✓

ALCO S/N: SCOTT08 ✓

### — BREATH ANALYSIS —

.000 ADJUSTED FOR 29.59 in ✓  
ALCO TARGET .079 12:21  
BLANK TEST .000 12:22  
INTERNAL STANDARD VERIFIED 12:22  
ALCO TV 29.59 in .000 ✓ 12:23  
BLANK TEST .000 12:24  
SUBJECT SAMPLE .000 12:24  
BLANK TEST .000 12:25  
ALCO TV 29.59 in .000 ✓ 12:25  
BLANK TEST .000 12:26

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130190

SEPTEMBER 18, 2009 ✓

TIME 12:20

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 29.59 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

### PRINTER TEST

! " # % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; ( = ) ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z I \ ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z ( ) \* +