

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

OCT 06 2009

Supervisor/Operator Performing the Verification Procedure: **DataMaster cdm S/N 130190** ✓

Name Bill Ferris ID# 5004 Date 09-28-09 <sup>27</sup> ~~10/14/09~~

A Agency Togiak Police Department Phone # 493-5212

Instrument Location Togiak Police Station 100 Main St. Togiak ✓

B Alco S/N 57023 ✓ Target Value .081 ✓ High Pressure 1000 PSI

Alco Test Values 0.87 ✓ 0.87 ✓  
1<sup>st</sup> Alco 2<sup>nd</sup> Alco

Signature Bill J. Ferris ✓

(OVER)

*BMB 10/13/09*

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 22<sup>nd</sup> day of Oct., 2009.

Carolyn M. Noland  
(Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130190 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130190

SEPTEMBER 27, 2009 ✓

OPERATOR'S NAME:  
FERRIS BILL J  
OPERATOR'S NUMBER: 5004  
SUBJECT'S LAST NAME:  
UIC

SUBJECT'S FIRST NAME/MI :  
XXXXXX

O.L. #: XX-XXX  
DEPT/AGENCY: T061  
CASE/REPORT: XX-XXXX  
TEST TYPE: U  
ALCO TARGET VALUE: 001 ✓  
ALCO S/N: 57023

### — BREATH ANALYSIS —

001 ADJUSTED FOR 30.64 in ✓  
ALCO TARGET .002 21:32  
BLANK TEST .000 21:33  
INTERNAL STANDARD VERIFIED 21:33  
ALCO TU 30.64 in .007 21:33  
BLANK TEST .000 21:34  
SUBJECT SAMPLE .000 21:35  
BLANK TEST .000 21:35  
ALCO TU 30.64 in .007 21:36  
BLANK TEST .000 21:36

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130190

SEPTEMBER 27, 2009 ✓  
TIME 21:29

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 30.64 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

! " # \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; ( = ) ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~ ¡ ¢