

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

MAR 17 2010

PCL XL error

Subsystem: KERNEL  
**Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program**

Error: IllegalTag

Operator: 0x1b **DataMaster cdm S/N** 130191

Supervisor/Operator Performing the Verification Procedure:  
Position: 18

Name BRANDON LAWRENSEN ID# 10609 Date 3/11/10

**A** Agency JUNEAU POLICE DEPT Phone # 586-0647

Instrument Location JUNEAU POLICE DEPT

Alco S/N X124646 Target Value .077 High Pressure 1000

**B** Alco Test Values .083 .082  
1<sup>st</sup> Alco 2<sup>nd</sup> Alco

Signature [Signature]

(OVER)

*BMB 3/10/2010*

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 5<sup>th</sup> day of April, 2010.

[Signature] (Notary Seal Stamp)

Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130191 ✓

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130191

MARCH 11, 2010 ✓

OPERATOR'S NAME: LAWRENSON/BRANDON/J  
OPERATOR'S NUMBER: 6609  
SUBJECT'S LAST NAME: VOC  
SUBJECT'S FIRST NAME/MI : VOC  
O.L. #: 0123456  
DEPT/AGENCY: JNU1  
CASE/REPORT: 0123456  
TEST TYPE: V ✓  
ALCO TARGET VALUE: .077 ✓  
ALCO S/N: X124646

--- BREATH ANALYSIS ---

.077 ADJUSTED FOR 29.02 in ✓  
ALCO TARGET .074 15:36  
BLANK TEST .000 15:37  
INTERNAL STANDARD VERIFIED 15:37  
ALCO TV 29.02 in .083 15:37  
BLANK TEST .000 15:38  
SUBJECT SAMPLE .000 15:38  
BLANK TEST .000 15:39  
BLANK TEST .000 15:40

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130191 ✓

MARCH 11, 2010 ✓  
TIME 15:32

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 50c  
BREATH TUBE: 43c  
BAROMETER: 29.02 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
CALIBRATION: OKAY

PRINTER TEST  
!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGH  
IJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnop  
qrstuvwxyz{|}~\*~