

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130191</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Robert J. Ely</u>	ID# <u>2681</u> Date <u>03-21-2008</u> ✓
Agency <u>Craig Police Department</u> ✓	Phone # <u>(907) 826-3330</u>
Instrument Location <u>506 2nd Street, Craig Alaska 99921</u>	
Alco S/N <u>X172967</u> ✓	Target Value <u>.081</u> ✓ High Pressure <u>400 PSI</u>
Alco Test Value Average <u>.084</u> 1 st Alco	<u>.085</u> 2 nd Alco
Signature <u>Robert J. Ely</u>	<u>CSB</u> <u>4/7/08</u>
(OVER)	

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Robert J Ely, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed

Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

4/12/08
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130191

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130191

MARCH 21, 2008 *COB 4/2/08*

OPERATOR'S NAME:
ELY, RJ
OPERATOR'S NUMBER: 2681
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/NI :
VER OF CAL
O.L. #: 123456789
DEPT/AGENCY: CR61
CASE/REPORT: 2008-XXXX
TEST TYPE: V
ALCO TARGET VALUE: .081
ALCO S/N: X172957

--- BREATH ANALYSIS ---

06/15/08

.081 ADJUSTED FOR 30.19 in		
ALCO TARGET	.081	12:14
BLANK TEST	.000	12:15
INTERNAL STANDARD	VERIFIED	12:15
ALCO TV 30.17 in	.084	12:16
BLANK TEST	.000	12:16
SUBJECT SAMPLE	.080 <i>COB 4/2/08</i>	12:17
BLANK TEST	.000	12:18
ALCO TV 30.19 in	.085	12:18
BLANK TEST	.000	12:19

10

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130191

MARCH 21, 2008
TIME 12:20

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 30.19 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcde{ghijklmnop
qrstuvwxyz{|}~`