

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130191 ✓
Name	Derek Sitz	ID#	4096 Date 3-27-09 ✓
A Agency	Anchorage Police Department	Phone #	764-9607
Instrument Location	Seely Substation ✓		
B Alco S/N	31085 ✓	Target Value	.080 ✓
		High Pressure	850 psi
Alco Test Values	.084 ✓		.083 ✓
	1 st Alco		2 nd Alco
Signature	Derek Sitz ✓		CSB 4/14/09

(OVER)

Under the Alaska Rules of Evidence, I certify that, (Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Derek Sitz, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz

Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

4/16/09
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130191 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130191

MARCH 27, 2009 ✓

OPERATOR'S NAME:
SITZ, DEREK W
OPERATOR'S NUMBER: 4096
SUBJECT'S LAST NAME:
SITZ
SUBJECT'S FIRST NAME/MI :
DEREK W
O.L. #: 1234567
DEPT/AGENCY: ANCI
CASE/REPORT: 00-1234
TEST TYPE: V
ALCO TARGET VALUE: .080 ✓
ALCO S/N: 31085

DS
1465
3-27-09

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 29.85 in		
ALCO TARGET	.079	23:50
BLANK TEST	.000	23:51
INTERNAL STANDARD	VERIFIED	23:51
ALCO TV 29.85 in	.084	23:51
BLANK TEST	.000	23:52
SUBJECT SAMPLE	.000	23:52
BLANK TEST	.000	23:53
ALCO TV 29.85 in	.083	23:53
BLANK TEST	.000	23:54

DS
1465
3-27-09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130191

MARCH 27, 2009 ✓
TIME 23:48

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 12/11/08
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.85 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

DS
1465
3-27-09

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~o