

VERIFICATION OF CALIBRATION REPORT APR 15 2009
of DataMaster cdm Breath Test Instrument
 State of Alaska APR 15 2009
 Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130191</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Patrick E. Martin</u>	ID# <u>4072</u> Date <u>4-15-09</u> ✓
A Agency <u>APD</u> Phone # <u>786-2601</u>	
Instrument Location <u>Seely Sub</u> ✓	
B Alco S/N <u>31085</u> ✓ Target Value <u>.080</u> ✓ High Pressure <u>550 psi</u>	
Alco Test Values	<u>.085</u> ✓ 1 st Alco
	<u>.084</u> ✓ 2 nd Alco
Signature <u>PE Martin</u> ✓	

08
5/13/09

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J Bolz
 Nita J Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 15th day of May, 2009.
Carolyn M. Noland (Notary Seal Stamp)

Carolyn M. Noland
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130191 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130191

APRIL 15, 2009 ✓

OPERATOR'S NAME:

MARTIN, PATRICK

OPERATOR'S NUMBER: 4072

SUBJECT'S LAST NAME:

MARTIN

SUBJECT'S FIRST NAME/MI :

PATRICK

O.L. #: 1234567

DEPT/AGENCY: ANCL

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .080 ✓

ALCO S/N: 31085 ✓

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 30.14 in ✓
ALCO TARGET .080 11:57
BLANK TEST .000 11:58
INTERNAL STANDARD VERIFIED 11:58
ALCO TV 30.14 in .085 ✓ 11:59
BLANK TEST .000 11:59
SUBJECT SAMPLE ^{MS} 5/13/09 .000 12:00
BLANK TEST .000 12:01
ALCO TV 30.14 in .084 ✓ 12:01
BLANK TEST .000 12:02

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130191

APRIL 15, 2009 ✓

TIME 11:54

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 12/11/08
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 30.14 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~Δ