

VERIFICATION OF CALIBRATION REPORT JUN 13 2008

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130191</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Robert J. Ely</u>	ID# <u>2681</u> ✓ Date <u>06-06-2008</u> ✓
Agency <u>Craig Police Department</u> ✓	Phone # <u>(907) 826-3330</u>
Instrument Location <u>506 2nd Street, Craig, Alaska 99921</u>	
Alco S/N <u>77286</u> ✓	Target Value <u>.079</u> ✓ High Pressure <u>1100</u> PSI
Alco Test Value Average <u>.084</u> ✓	<u>.084</u> ✓
1 st Alco	2 nd Alco
Signature <u><i>Robert J. Ely</i></u>	<u>CSB</u> <u>6/27/08</u>
(OVER)	


(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Robert J Ely, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

6/30/08
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130191

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130191

JUNE 06, 2008

OPERATOR'S NAME:

ELY, RJ

OPERATOR'S NUMBER: 2681

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

VER OF CAL NEW BOTTLE

O.L. #: 123456789

DEPT/AGENCY: CR61

CASE/REPORT: 2008-XXXX

TEST TYPE: V

ALCO TARGET VALUE: .079 ✓

ALCO S/N: 77286 ✓

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.19 in
ALCO TARGET .079 20:58
BLANK TEST .000 20:58
INTERNAL STANDARD VERIFIED 20:58
ALCO TV 30.19 in .004 ✓ 20:59
BLANK TEST .000 21:00
SUBJECT SAMPLE .000 21:00
BLANK TEST .000 21:01
ALCO TV 30.19 in .004 21:01
BLANK TEST .000 21:02

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130191

JUNE 06, 2008 ✓

TIME 20:53

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 30.19 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~