


JUN 27 2011

VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument
State of Alaska*

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130192</u>
Name	<u>TRACY MURPHY</u>	ID#	<u>7752</u> Date <u>6-6-11</u>
A Agency	<u>JUNEAU POLICE DEPT</u>	Phone #	<u>907-957-0600</u>
Instrument Location		<u>JUNEAU POLICE DEPT.</u>	
Alco S/N		Target Value	High Pressure
<u>X301550</u>		<u>.078</u>	<u>30.48</u>
Alco Test Values		1 st Alco	2 nd Alco
<u>.080</u>		<u>.080</u>	
Signature			


(OVER)

*DMB
6/27/11*

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 22 day of July, 2011.

 (Notary Seal Stamp)
 Daniel L. Nicholas Nikki Roth
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130192

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130192

JUNE 06, 2011

OPERATOR'S NAME:
MURPHY/TRACY/E
OPERATOR'S NUMBER: 7252
SUBJECT'S LAST NAME:
MURPHY
SUBJECT'S FIRST NAME/MI :
TRACY
O.L. #: VOC
DEPT/AGENCY: JNU1
CASE/REPORT: VOC
TEST TYPE: V
ALCO TARGET VALUE: .078
ALCO S/N: X301550

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 30.48 in
ALCO TARGET .079 07:56
BLANK TEST .000 07:57
INTERNAL STANDARD VERIFIED 07:57
ALCO TV 30.48 in .080 07:57
BLANK TEST .000 07:58
SUBJECT SAMPLE .000 07:59
BLANK TEST .000 07:59
ALCO TV 30.48 in .080 08:00
BLANK TEST .000 08:00

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130192

JUNE 06, 2011
TIME 07:54

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 30.48 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnop
qrstuvwxyz{|}~

