

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument


State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130192</u> ✓
Name	<u>BRANDON LAWRENSON</u>	ID# <u>6609</u> Date <u>01/22/09</u> ✓
Agency	<u>JUNEAU POLICE DEPT</u>	Phone # <u>586-0600</u>
Instrument Location	<u>JUNEAU POLICE DEPT</u>	
Alco S/N	<u>X124623</u> ✓	Target Value <u>.081</u> ✓ High Pressure <u>500</u>
Alco Test Values	<u>.086</u> ✓ 1 st Alco	<u>.085</u> ✓ 2 nd Alco
Signature		
(OVER)		

Under the Alaska Rules of Evidence, I certify that, (Do Not write in the area below)

- (a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.
- (b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Brandon Lawrenson, whose name appears in my records, is qualified to verify the calibration of the above instrument.
- (c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,



Jeanne Swartz
Scientific Director, Designee
State of Alaska
State Breath Alcohol Testing Program

2/9/09

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130192 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130192

JANUARY 22, 2009 ✓

OPERATOR'S NAME:
LAWRENSEN/BRANDON/J
OPERATOR'S NUMBER: 6609
SUBJECT'S LAST NAME:

VOC
SUBJECT'S FIRST NAME/MI :
VOC

O.L. #: 1234567
DEPT/AGENCY: JNU1
CASE/REPORT: 012345
TEST TYPE: V
ALCO TARGET VALUE: .081 ✓
ALCO S/N: X124623

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 30.74 in
ALCO TARGET .083 04:52
BLANK TEST .000 04:53
INTERNAL STANDARD VERIFIED 04:53
ALCO TV 30.74 in .086 ✓ 04:53
~~BLANK TEST~~ .000 04:54
SUBJECT SAMPLE .000 04:55
BLANK TEST .000 04:55
ALCO TV 30.74 in .085 ✓ 04:56
BLANK TEST .000 04:57

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130192

JANUARY 22, 2009 ✓
TIME 04:51

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 43c
BAROMETER: 30.74 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGH
IJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnop
qrstuvwxyz{|}~*~