VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130192

Supervisor/Operator Performing the Verification Procedure:

A

Name STEVE CHRISTENSEN ID# 1315 Date 4/21/08

Agency JUNEAU Police Dept Phone # 907-586-0600

Instrument Location JUNEAU Police Dept

B

Alco S/N 70108 ✓ Target Value .080 ✓ High Pressure 500

Alco Test Values .081 (1st Alco) .081 (2nd Alco)

Signature [signature] CDB 4/28/08

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Steve Christensen , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

[signature] 4/29/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

BT9 06/02

APR 24 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program

DataMaster cdm S/N 130192

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130192

APRIL 21, 2008 COB 4/28/08

OPERATOR'S NAME:
CHRISTENSEN S,E
OPERATOR'S NUMBER: 1315
SUBJECT'S LAST NAME:
VERFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: 042108
DEPT/AGENCY: JNU1
CASE/REPORT: 042108
TEST TYPE: V
ALCO TARGET VALUE: .080
ALCO S/N: 70108

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 30.48 in

ALCO TARGET	.081	19:32
BLANK TEST	.000	19:33
INTERNAL STANDARD	VERIFIED	19:33
ALCO TV 30.48 in	.081	19:34
BLANK TEST	.000	19:35
SUBJECT SAMPLE	.000	19:35
BLANK TEST	.000	19:36
ALCO TV 30.48 in	.081	19:36
BLANK TEST	.000	19:37

COB 4/28/08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130192

APRIL 21, 2008
TIME 19:37

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	30.48 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~