


VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130192</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>BRANDON LAWRENSON</u>	ID# <u>66699</u> Date <u>07/15/08</u> ✓
A Agency <u>JUNEAU POLICE DEPT</u>	Phone # <u>586-0600</u>
Instrument Location <u>JUNEAU PD</u> ✓	
B Alco S/N <u>70108</u> ✓ Target Value <u>.080</u> ✓ High Pressure <u>100</u>	
Alco Test Values	<u>.081</u> ✓ <u>.081</u> ✓ 1 st Alco 2 nd Alco
Signature <u>[Signature]</u>	<u>CSB</u> <u>7/22/08</u>
(OVER)	

Under the Alaska Rules of Evidence, I certify that, (Do Not write in the area below)

- (a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.
- (b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Brandon Lawrenson , whose name appears in my records, is qualified to verify the calibration of the above instrument.
- (c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

[Signature]

Jeanne Swartz
 Scientific Director Designee
 State of Alaska
 State Breath Alcohol Testing Program

7/22/08
 Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130192

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130192

JULY 15, 2008 *7/15/08*

OPERATOR'S NAME:

LAWRENSON/BRANDON/J

OPERATOR'S NUMBER: 6609

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI:

UOC

D.L.#: 0000000

DEPT/AGENCY: JUNI

CASE/REPORT: 012345

TEST TYPE: U

ALCO TARGET VALUE: *0.081*

ALCO S/N: 70100

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.43 in
ALCO TARGET .081 14:07
BLANK TEST .000 14:08
INTERNAL STANDARD VERIFIED 14:08
ALCO TU 30.43 in .081 14:08
BLANK TEST .000 14:09
SUBJECT SAMPLE .000 *0.081* 14:10
BLANK TEST .000 14:11
ALCO TU 30.43 in .081 14:11
BLANK TEST .000 14:12

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130192

JULY 15, 2008

TIME: 14:03

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 30.43 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLmnopqrstuvwxyz[\]^_`abcdefgijklmnop
pqrstuvwxy{|}~