

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130192 ✓
Name	BRANDON LAWRENSEN	ID#	6609 Date 08/30/08 ✓
Agency	JUNEAU POLICE DEPT	Phone #	586-0600
Instrument Location	JUNEAU POLICE DEPT ✓		
Alco S/N		Target Value	High Pressure
X124623 ✓		.081 ✓	900 PSI
Alco Test Values		1 st Alco	2 nd Alco
		.084 ✓	.084 ✓
Signature			083 9/12/08
(OVER)			


Under the Alaska Rules of Evidence, I certify that, (Do Not write in the area below)


(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Brandon Lawrenson, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,





Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

9/15/08

Date

BT9 03:00

SEP 04 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130192 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130192

AUGUST 30, 2008 ✓ *COB 9/12/08*

OPERATOR'S NAME:
LAWRENSEN/BRANDON/J
OPERATOR'S NUMBER: 6609
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: 1234567
DEPT/AGENCY: JNU1
CASE/REPORT: 012345
TEST TYPE: V
ALCO TARGET VALUE: .081 ✓
ALCO S/N: X124623 ✓

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 30.29 in		
ALCO TARGET	.082	15:09
BLANK TEST	.000	15:10
INTERNAL STANDARD	VERIFIED	15:10
ALCO TV 30.29 in	.084 ✓	15:10
BLANK TEST	.000	15:11
SUBJECT SAMPLE	.000 <i>COB 9/12/08</i>	15:12
BLANK TEST	.000	15:13
ALCO TV 30.29 in	.084 ✓	15:13
BLANK TEST	.000	15:14

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130192

AUGUST 30, 2008 ✓
TIME 15:07

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS:	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	30.29 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnop
qrstuvwxyz{|}~