

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N ~~130191~~

(CONTINUED FROM FRONT PAGE)

130192

NJB
12/17/10

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130192

OCTOBER 02, 2010

OPERATOR'S NAME:
MURPHY/TRACY/ERNEST
OPERATOR'S NUMBER: 7252
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: VOC
DEPT/AGENCY: JNU1
CASE/REPORT: VOC
TEST TYPE: V
ALCO TARGET VALUE: .079
ALCO S/N: 77621

BMB
12/17/10

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.27 in
ALCO TARGET .079 17:32
BLANK TEST .000 17:33
INTERNAL STANDARD VERIFIED 17:33
ALCO TV 30.27 in .081 17:33
BLANK TEST .000 17:34
SUBJECT SAMPLE .000 17:34
BLANK TEST .000 17:35
ALCO TV 30.27 in .080 17:35
BLANK TEST .000 17:36

BMB
12/17/10

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130192

OCTOBER 02, 2010
TIME 17:28

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 30.27 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABC
HIJKLMNPOQRSTUVWXYZ[\]^_`abcdefgijkl
pqrstuvwxyz{|}~