VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130193

Supervisor/Operator Performing the Verification Procedure:

A

Name Kyle Carrington ID# 5336 Date 8-4-08

Agency University of Alaska Fairbanks Police Dept. Phone # 474-7721

Instrument Location 612 Yukon Drive Fairbanks, AK 99775

B

Alco S/N X124658 Target Value .082 High Pressure 950 PSI

Alco Test Values .082 (1st Alco) .082 (2nd Alco)

Signature Kyle S. Carrington CDB 8/12/08

(OVER)

Under the Alaska Rules of Evidence, I certify that: (Do not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Kyle Carrington, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

Jeanne Swartz 8/13/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

AUG 07 2008

BT9 06/03

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130193

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130193

AUGUST 04, 2008 CAB 8/19/08

OPERATOR'S NAME:
CARRINGTON,KYLE
OPERATOR'S NUMBER: 5336
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
I
O.L. #: 0
DEPT/AGENCY: FBK7
CASE/REPORT: 0
TEST TYPE: U
ALCO TARGET VALUE: .082
ALCO S/N: X124658

— BREATH ANALYSIS —

.082 ADJUSTED FOR 29.60 in

ALCO TARGET	.081	07:38
BLANK TEST	.000	07:39
INTERNAL STANDARD	VERIFIED	07:39
ALCO TV 29.60 in	.082	07:39
BLANK TEST	.000	07:40
SUBJECT SAMPLE	.000	07:41
BLANK TEST	.000	07:42
ALCO TV 29.60 in	.082	07:42
BLANK TEST	.000	07:43

CAB 8/19/08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130193

AUGUST 04, 2008
TIME 07:36

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.60 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~