VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130193

Supervisor/Operator Performing the Verification Procedure:

A

Name Colleen O'Bryant ID# 6237 Date 12/11/09

Agency SCDL Phone # 269-5740

Instrument Location Palmer PD

B

Alco S/N X124671 Target Value 0.084 High Pressure 1100

Alco Test Values 0.085 (1st Alco) 0.084 (2nd Alco)

Signature C. O'Bryant CB 1/25/10

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 24th day of February, ~~2009~~ 2010.

(Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

2/24/10

BT9 06/03

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130193

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130193

DECEMBER 11, 2009 COB 1/25/10

OPERATOR'S NAME:
OBRYANT/COLLEEN/S
OPERATOR'S NUMBER: 6237
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: 012345
DEPT/AGENCY: AAA1
CASE/REPORT: 09-1234
TEST TYPE: V
ALCO TARGET VALUE: ~~.076~~ 0.084
ALCO S/N: X124671 COB 12/1/09

--- BREATH ANALYSIS ---

.084 ADJUSTED FOR 29.94 in

ALCO TARGET	.084	18:29
BLANK TEST	.000	18:30
INTERNAL STANDARD	VERIFIED	18:31
ALCO TV 29.94 in	.085	18:31
BLANK TEST	.000	18:32
SUBJECT SAMPLE COB 1/25/10	.000	18:32
BLANK TEST	.000	18:33
ALCO TV 29.94 in	.084	18:33
BLANK TEST	.000	18:34

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130193

DECEMBER 11, 2009
TIME 18:22

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.94 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~