

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130194</u>
Name	<u>Jaime Smith</u>	ID# <u>6762</u> Date <u>5/31/12</u>
A Agency	<u>Denali National Park</u>	Phone # <u>907-683-9575</u>
Instrument Location <u>Denali NP Headquarters</u>		
B Alco S/N	<u>4M</u>	Target Value <u>0.080</u> High Pressure <u>750 psi</u>
Alco Test Values		
	<u>0.075</u> 1 <sup>st</sup> Alco	<u>0.074</u> 2 <sup>nd</sup> Alco
Signature <u>Jaime Smith</u>		

(OVER)

BMBS  
5/14/12

(Do Not write in the area below)



I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 3rd day of July, 2012.  
NAK 7/3/12

Nikki Roth (Notary Seal Stamp)  
Carolyn M. Notland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130194

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130194

MAY 31, 2012

OPERATOR'S NAME:

SMITH, JAIME L

OPERATOR'S NUMBER: 6752

SUBJECT'S LAST NAME:

X

SUBJECT'S FIRST NAME/MI :

X

O.L. #: X

DEPT/AGENCY: DNL3

CASE/REPORT:

TEST TYPE: V

ALCO TARGET VALUE: 0.350

ALCO S/N: 4M

### --- BREATH ANALYSIS ---

.000 ADJUSTED FOR 27.62 in  
ALCO TARGET .073 12:55  
BLANK TEST .000 12:56  
INTERNAL STANDARD VERIFIED 12:56  
ALCO TV 27.62 in .075 12:57  
BLANK TEST .000 12:57  
SUBJECT SAMPLE .000 12:58  
BLANK TEST .000 12:58  
ALCO TV 27.62 in .074 12:59  
BLANK TEST .000 12:59

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130194

MAY 31, 2012

TIME 13:01

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 30c  
BREATH TUBE: 41c  
BAROMETER: 27.62 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!@#\$%^&\*~.-/0123456789:;<=>?@BCDEFG  
HIJKL MNOPQRSTUVWXYZ-1\_`abcde fghij klmno  
pqrstuvwxyz{|}~+

