

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130194 ✓
Name		Lucas P Amidon	ID# 5657 Date 4/18/08 ✓
A	Agency	BRW North Slope Borough PD	Phone # 907 852 6111
Instrument Location		Barrow	
Alco S/N		X 172978 ✓	Target Value .083 ✓ High Pressure 495
Alco Test Values		.065	.065
		1 st Alco	2 nd Alco
Signature		Lucas P Amidon	COB 4/22/08
(OVER)			

Under the Alaska Rules of Evidence, I certify that, (Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Lucas Amidon, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

J Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

4/27/08
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130194

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130194

APRIL 18, 2008

OPERATOR'S NAME:
ANTHON-LUCAS/P

OPERATOR'S NUMBER: 5651

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

FILE #: 4182008

DEPT/AGENCY: BRWI

CASE/REPORT: 04182008

TEST TYPE: V

ALCO TARGET VALUE: .083

ALCO S/N: X172978

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 30.51 in

ALCO TARGET	.084	13:23
BLANK TEST	.000	13:23
INTERNAL STANDARD	VERIFIED	13:23
ALCO @ 30.51 in	.085	13:24
BLANK TEST	.000	13:25
SUBJECT SAMPLE	.083	13:25
BLANK TEST	.000	13:26
ALCO @ 30.50 in	.085	13:26
BLANK TEST	.000	13:27

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130194

APRIL 18, 2008

TIME 13:20

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 40c

BAROMETER: 30.51 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ*+,-.:/\|_`~!@#\$%^&*
(){}~!@#\$%^&*(){}~!@#\$%^&*()