

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>1301941</u>	
Name <u>JAMIES MICHELS</u>		ID# <u>6502</u>	Date <u>6/26/08</u> ✓
A	Agency <u>NORTH SLOPE BOROUGH POLICE DEPT</u> ✓	Phone # <u>852-6111</u>	
Instrument Location <u>SQUAD ROOM</u>			
Alco S/N <u>X124617-8</u> ✓ <u>7/7/08</u>		Target Value <u>.079</u> ✓	High Pressure <u>900</u> <u>pse</u>
B	Alco Test Values <u>.078</u> ✓		<u>.078</u> ✓
	1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco	
Signature <u>James Michels</u>			<u>CSB</u> <u>7/3/08</u>
(OVER)			


Under the Alaska Rules of Evidence, I certify that (Do not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, James Michels, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed



Jeanne Swartz

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

7/7/08  
Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130194

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130194

JUNE 26, 2008

OPERATOR'S NAME:

NICHOLS/JAMES/D

OPERATOR'S NUMBER: 6582

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: VOC

DEPT/AGENCY: DMW

BASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .079 ✓

ALCO S/N: X124617 ✓

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.94 in  
ALCO TARGET .079 09:35  
BLANK TEST .000 09:35  
INTERVAL STANDARD VERIFIED 09:36  
ALCO @ 29.94 in .078 ✓ 09:36  
BLANK TEST .000 09:37  
SUBJECT SAMPLE .000 09:37  
BLANK TEST .000 09:38  
ALCO @ 29.94 in .078 ✓ 09:38  
BLANK TEST .000 09:39

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130194

JUNE 26, 2008 ✓

TIME 09:32

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 42c  
BREATH TUBE: 42c  
BAROMETER: 29.94 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF  
HIJKLMN.OPQRSTUWXYZ[\]^\_`abcde fghijklmno  
pqrstuvwxyz{|}~"