

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130 194 ✓
Name		Lucas P Amidon	ID# 5651 Date 10/14/08 ✓
A	Agency	North Slope Borough PD	Phone # 852-6111
Instrument Location		Barrow ✓	
Alco S/N		X124617 ✓	Target Value .079 ✓ High Pressure 550 ✓
Alco Test Values		.079 ✓ 1 st Alco	.079 ✓ 2 nd Alco
Signature		Lucas P Amidon ✓	CSB 10/24/08
(OVER)			

Under the Alaska Rules of Evidence, I certify that, ^(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Lucas Amidon, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed



Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

10/24/08
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130194 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130194

OCTOBER 14, 2003 ✓
10/24/03

OPERATOR'S NAME:
ANDON/LUCAS/P
OPERATOR'S NUMBER: 5651
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: 10142008
DEPT/AGENCY: BRW1
CASE/REPORT: 10142008
TEST TYPE: V
ALCO TARGET VALUE: .079 ✓
ALCO S/N: X124617 ✓

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.17 in
ALCO TARGET .079 12:55
BLANK TEST .000 12:55
INTERNAL STANDARD VERIFIED 12:55
ALCO TV 30.17 in .079 ✓ 12:56
BLANK TEST .000 12:57
SUBJECT SAMPLE .000 *10/24/03* 12:57
BLANK TEST .000 12:58
ALCO TV 30.17 in .078 ✓ 12:58
BLANK TEST .000 12:59

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130194

OCTOBER 14, 2003 ✓
TIME 12:52

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHANGER: 49c
BREATH TUBE: 42c
BAROMETER: 30.17 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~