

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130194 12-19-08
CMM

Name JAMES MICHELS ID# 6502 Date 12-16-08

A Agency USBPID Phone # 852-6111

Instrument Location SQUAD ROOM - BARROW

B Alco S/N X124617 ✓ Target Value .079 ✓ High Pressure 200 psf


Alco Test Values .080 ✓ .080 ✓
1st Alco 2nd Alco

Signature SGT James Michels ✓ CMB
12/23/08

(OVER)

Under the Alaska Rules of Evidence, I certify that: (Do Not write in the area below)

- (a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.
- (b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, James Michels, whose name appears in my records, is qualified to verify the calibration of the above instrument.
- (c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed

Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

12/23/08

Date

DEC 18 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130194 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130194

DECEMBER 15, 2003 ✓
CBS 12/23/03

OPERATOR'S NAME:

NICHOLS/JAMES/D

OPERATOR'S NUMBER: 6522

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI:

VOC

D.L.I. #: VOC

DEPT/AGENCY: BROW

CRIME/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .079 ✓

ALCO S/N: X124617 ✓

--- BREATH ANALYSIS ---

ALCO ADJUSTED FOR 30.43 in

ALCO TEST	.000	20:13
BLANK TEST	.000	20:13
INTERNAL STANDARD	VERIFIED	22:13
ALCO TO 30.43 in	.000 ✓	20:14
BLANK TEST	.000	20:14
SUBJECT SAMPLE	.000	20:13
BLANK TEST	.000	20:16
ALCO TO 30.45 in	.000 ✓	20:16
BLANK TEST	.000	20:17

CBS 12/23/03

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130194

DECEMBER 15, 2003 ✓
TEST TIME

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

PROJECTED: 30.43 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

ALCO TEST: 130194/01234567891011121314151617181920
ALCO TEST: 130194/01234567891011121314151617181920
ALCO TEST: 130194/01234567891011121314151617181920