

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130194 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130194

APRIL 18, 2010 ✓ *COB 4/20/10*

OPERATOR'S NAME:

MCCARRIGAN DEREK A

OPERATOR'S NUMBER: 5420

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 1234567

DEPT/AGENCY: KET1

CASE/REPORT: 10-2

TEST TYPE: V

ALCO TARGET VALUE: .084 ✓

ALCO S/N: 57447

--- BREATH ANALYSIS ---

.084 ADJUSTED FOR 30.22 in ✓
ALCO TARGET .084 07:56
BLANK TEST .000 07:57
INTERNAL STANDARD VERIFIED 07:57
ALCO TV 30.22 in .091 ✓ 07:57
BLANK TEST .000 07:58
SUBJECT SAMPLE *COB 4/20/10* .000 07:58
BLANK TEST .000 07:59
ALCO TV 30.22 in .091 07:59
BLANK TEST .000 08:00

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130194

APRIL 18, 2010 ✓

TIME 07:53

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 30.22 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrs tuvwxyz{|}~

VERIFICATION OF CALIBRATION REPORT

DataMaster cdm Breath Test Instrument

State of Alaska

APR 21 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130194

Name Derek A. McGARRIGAN ID# 5420 Date 4/18/10

A Agency Ketchikan Police Dept. Phone # 225-6631

Instrument Location Ketchikan Police Dept, 361 Main St., Ketchikan, AK 99901

B Alco S/N 57447 Target Value .084 High Pressure 800

Alco Test Value Average .091 .091

1st Alco 2nd Alco

Signature [Signature]

(OVER)

*CSB
4/18 CSB
4/20/10*

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 20th day of April, 2010.

[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

