

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUN 07 2010

Supervisor/Operator Performing the Verification Procedure:		
DataMaster cdm S/N <u>130 194</u> ✓		
Name <u>Trooper James Kimura</u>	ID# <u>5677</u> Date <u>5/29/10</u> ✓	
A Agency <u>Alaska State Troopers</u> Phone # <u>(907) 225-5118</u>		
Instrument Location <u>Ketchikan Police Department (KPD)</u> ✓		
B Alco S/N <u>X301765</u> ✓ Target Value <u>.079</u> ✓ High Pressure <u>1,000</u>		
Alco Test Values	<u>-.084</u> ✓ 1 <sup>st</sup> Alco	<u>-.084</u> ✓ 2 <sup>nd</sup> Alco
Signature <u>[Signature]</u> ✓		<u>CVB</u> <u>6/11/10</u>
(OVER)		

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 24<sup>th</sup> day of June, 2010.

[Signature]  
(Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130 194 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130194 ✓

MAY 29, 2010 ✓  
TIME 20:01

*COB  
9/1/10*

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130194

MAY 29, 2010

OPERATOR'S NAME:  
KIMURA/JAMES/K  
OPERATOR'S NUMBER: 5677  
SUBJECT'S LAST NAME:  
VOC  
SUBJECT'S FIRST NAME/MI :  
VOC  
O.L. #: 000000  
DEPT/AGENCY: KETA  
CASE/REPORT: 10-000033  
TEST TYPE: V ✓  
ALCO TARGET VALUE: .079 ✓  
ALCO S/N: X301745 ✓

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.56 in ✓  
ALCO TARGET .000 20:05  
BLANK TEST .000 20:06  
INTERNAL STANDARD VERIFIED 22:06  
ALCO TV 30.56 in .004 ✓ 20:07  
BLANK TEST .000 20:07  
SUBJECT SAMPLE *COB  
9/1/10* .000 20:08  
BLANK TEST .000 20:08  
ALCO TV 30.56 in .004 ✓ 20:09  
BLANK TEST .000 20:09

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130194 ✓

MAY 29, 2010 ✓  
TIME 19:50

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 30.56 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF  
GHIJKLMNOPQRSTUVWXYZ\*\_`abcde fghijklmno  
pqrstuvwxyz{|}~