

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130194 ✓

Name Jaime Smith ID# 6762 ✓ Date 11/08/2010 ✓

A Agency Denali National Park Phone # 907-683-9525

Instrument Location Denali National Park - Headquarters

---

B Alco S/N X301591 ✓ Target Value 0.080 ✓ High Pressure 1250

Alco Test Values .079 ✓ .078 ✓

1<sup>st</sup> Alco 2<sup>nd</sup> Alco

Signature Jaime Smith

(OVER)

*DMB*  
*12/16/10*

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

*Nita J. Bolz*  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 23<sup>RD</sup> day of DEC, 2010.

*Carolyn M. Noland* (Notary Seal Stamp)  
 Carolyn M. Noland DANIEL NICHOLAS  
 Notary Public, State of Alaska  
 Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130194

### Breath Test Record

Serial Number 130194 Model DataMaster  
 Location BRW1 North Slope Borough Police Dept.

(IT PAGE)

TO SUBJECT TEST AND THE DIAGNOSTIC CHECKS.

### Subject Information

Last Name Verification Of Cal O.L. # X  
 First Name X

### Operator Information

Operator Name Smith/J Case   
 Operator # 6762 Department DNLB

### DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
 DATA MASTER cdm 130194

11 NOVEMBER 20, 2010  
 TIME 14:52

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
 PROGRAM: OKAY  
 SOFTWARE DATE: 10/10/06  
 HEATER: OKAY  
 SAMPLE CHAMBER: 50c  
 BREATH TUBE: 40c  
 BAROMETER: 27.83 in  
 FLOW RATE TUB: OKAY  
 PUMP: OKAY  
 HIGH PRESS: OKAY  
 DETECTOR: OKAY  
 FILTERS: OKAY  
 CHECKS STANDARD: OKAY  
 CALIBRATION: OKAY

PRINTER TEST  
 Test 1: 11/20/10 14:52:11 (E) 0.000000  
 Test 2: 11/20/10 14:52:12 (E) 0.000000  
 Test 3: 11/20/10 14:52:13 (E) 0.000000

### Test Results

Alco Target 0.080 Alco Serial X301591  
 Nominal 0.074 Barometer 27.82 Test Type V  
 Dry Gas 1 0.079 Time: 14:47:11 Test Date 11/08/2010  
 Subject Alcohol 0.000 Time: 14:48:28 Test Time 14:45:40  
 Dry Gas 2 0.078 Time: 14:49:27 Report Y

DATA MASTER cdm 130194  
 SERIAL OF FIRST VERIFICATION :  
 O.L. # :  
 AGENCY: DNLB  
 CASE REPORT:  
 TEST TYPE: V  
 ALCO TARGET VALUE: 0.080  
 ALCO SERIAL: X301591

--- BREATH ANALYSIS ---

TEST INITIATED FOR 27.83 in  
 ALCO TEST: .074 14:45  
 BLANK TEST: .000 14:46  
 INTERNAL FLOWMETER: 27.82 14:46  
 DRY GAS 1: .079 14:47  
 DRY GAS 2: .078 14:47  
 SUBJECT ALCOHOL: .000 14:48  
 DRY GAS 1: .079 14:49  
 DRY GAS 2: .078 14:49  
 DRY GAS TEST: .000 14:50

*Handwritten:* DMS 12/10/10