

VERIFICATION OF CALIBRATION REPORT AUG 31 2011
of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N <u>130194</u>					
Name <u>Jaime Smith</u>	ID# <u>6762</u> Date <u>8/26/11</u> BMB 9/2/11				
A Agency <u>Denali National Park</u> Phone # <u>907-683-9525</u>					
Instrument Location <u>Denali National Park - Headquarters</u>					
B Alco S/N <u>X301591</u> Target Value <u>0.080</u> High Pressure <u>1150 PSI</u>					
Alco Test Values	<table style="width: 100%; border: none;"> <tr> <td style="text-align: center; border-bottom: 1px solid black;"><u>.078</u></td> <td style="text-align: center; border-bottom: 1px solid black;"><u>.079</u></td> </tr> <tr> <td style="text-align: center; font-size: small;">1st Alco</td> <td style="text-align: center; font-size: small;">2nd Alco</td> </tr> </table>	<u>.078</u>	<u>.079</u>	1 st Alco	2 nd Alco
<u>.078</u>	<u>.079</u>				
1 st Alco	2 nd Alco				
Signature <u>Jaime Smith</u>					

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 2 day of Sept, 2011. AK

Nikki Roth (Notary Seal Stamp)
 Carolyn M. Noland Nikki Roth
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130194

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATA MASTER cdm 130194

AUGUST 26, 2011

OPERATOR'S NAME:
SMITH/JAMEL

INSTRUMENT'S NUMBER: 6752

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

X

O.L. #: X

DEPT/AGENCY: DMLB

CASE/REPORT:

TEST TYPE: V

ALCO TARGET VALUE: .000

ALCO S/N: K301531

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 27.95 in

ALCO TARGET	.074	13:59
BLANK TEST	.000	14:01
INTERNAL STANDARD	VERIFIED	14:01
ALCO @ 27.95 in	.078	14:01
BLANK TEST	.000	14:02
SUBJECT SAMPLE	.000	14:02
BLANK TEST	.000	14:03
ALCO @ 27.95 in	.079	14:03
BLANK TEST	.000	14:04



DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATA MASTER cdm 130194

AUGUST 26, 2011

TIME 14:05

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	53c
BREATH TUBE:	42c
BAROMETER:	27.95 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

ALCO TARGET VALUE: .074
INTERNAL STANDARD VERIFIED
SUBJECT SAMPLE