VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130194.

Supervisor/Operator Performing the Verification Procedure:

A

Name John Fish. ID# 7674 Date 12-20-11.

Agency UPS Phone # 907 683-7587.

Instrument Location Dowl. UPS.

B

Alco S/N X301591 Target Value .080 High Pressure 1000.

Alco Test Values .078 (1st Alco) .078 (2nd Alco)

Signature

(OVER)

BMB 1/13/12

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 2nd day of Feb, ~~2010~~ 2012.

Nikki Roth (Notary Seal Stamp)
~~Carolyn M. Noland~~ Nikki Roth
Notary Public, State of Alaska
Commission Expires with Office

NIKKI ROTH
My Commission Expires
NOTARY PUBLIC
With Office
STATE OF ALASKA

STATE OF ALASKA
SCIENTIFIC CRIME DETECTION LABORATORY

BT9 06/03

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130194

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130194

DECEMBER 20, 2011

BMB 1/13/12

OPERATOR'S NAME:
FISH/JOHN/P
OPERATOR'S NUMBER: 7674
SUBJECT'S LAST NAME:
X
SUBJECT'S FIRST NAME/MI :
X
O.L. #: X
DEPT/AGENCY: DNL3
CASE/REPORT:
TEST TYPE: V
ALCO TARGET VALUE: .080
ALCO S/N: X301591

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 27.85 in

ALCO TARGET	.074	12:36
BLANK TEST	.000	12:37
INTERNAL STANDARD	VERIFIED	12:37
ALCO TV 27.83 in	.078	12:38
BLANK TEST	.000	12:39
SUBJECT SAMPLE	.000	12:39
BLANK TEST	.000	12:40
ALCO TV 27.83 in	.078	12:40
BLANK TEST	.000	12:41

BMB 1/13/12

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130194

DECEMBER 20, 2011
TIME 12:46

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 43.7
BREATH TUBE: 41.0
BAROMETER: 27.85 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~