

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

JUL 27 2011

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130195

Supervisor/Operator Performing the Verification Procedure:

A Name Robert J. Ely ID# 2681 Date 07-21-2011

Agency Craig Police Department Phone # (907) 826-3330

Instrument Location 506 2nd Street, Craig, Alaska 99921

B Alco S/N 75509 Target Value .079 High Pressure 600 PSI

Alco Test Value Average .081 .080
1st Alco 2nd Alco

Signature Robert J. Ely

(OVER)

(Do Not write in the area below)

BMB
7/27/11

I, Nita J. Bolz, after being first duly sworn, depose and state as follows;

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 12 day of Aug, 2010

NJB
2011

Nikki Roth (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



