VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

SEP 2 6 2011

DataMaster cdm S/N 130195

Supervisor/Operator Performing the Verification Procedure:

A

Name Robert J. Ely ID# 2681 Date 09-22-2011

Agency Craig Police Department Phone # (907) 826-3330

Instrument Location 506 2nd Street, Craig Alaska 99921

B

Alco S/N 75509 Target Value .079 High Pressure 550 PSI

Alco Test Value Average .080 (1st Alco) .079 (2nd Alco)

Signature Robert J. Ely

CBB 9/26/11

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows;

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons Responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 7 day of Nov, 2011 ~~2010~~

Nikki Roth (Notary Seal Stamp)
~~Daniel L. Nicholas~~ Nikki Roth
Notary Public, State of Alaska
Commission Expires with Office

NIKKI ROTH
My Commission Expires
NOTARY PUBLIC
With Office
STATE OF ALASKA

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130195

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130195

SEPTEMBER 22, 2011

OPERATOR'S NAME: CB 9/26/11
ELY, RJ
OPERATOR'S NUMBER: 2681
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
VER OF CAL
D.L. #: 123456789
DEPT/AGENCY: CRS1
CASE/REPORT: 11-XXXXXX
TEST TYPE: V
ALCO TARGET VALUE: .079
ALCO S/N: 75509

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.88 in

ALCO TARGET	.078	10:39
BLANK TEST	.000	10:40
INTERNAL STANDARD	VERIFIED	10:40
ALCO TV 29.88 in	.080	10:41
BLANK TEST	.000	10:41
SUBJECT SAMPLE	.000	10:42
[illegible] TEST	.000	10:43
ALCO TV 29.88 in	.079	10:43
BLANK TEST	.000	10:44

CB 9/26/11

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130195

SEPTEMBER 22, 2011
TIME 10:37

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.88 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~