VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130195

Supervisor/Operator Performing the Verification Procedure:

A

Name Robert J. Ely ID# 2681 Date 04-09-2012

Agency Craig Police Department Phone # (907) 826-3330

Instrument Location 506 2nd Street, Craig Alaska 99921

B

Alco S/N 75509 Target Value .079 High Pressure 350 PSI

Alco Test Value Average .082 (1st Alco) .081 (2nd Alco)

Signature Robert J. Ely

(OVER)

BNB 4/12/12

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows;

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons Responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 3rd day of July, ~~2010~~ 2012 NAK 7/3/12

Nikki Roth (Notary Seal Stamp)

NAK ~~Daniel L. Nicholas~~ Nikki Roth 7/3/12
Notary Public, State of Alaska
Commission Expires with Office

NIKKI ROTH
My Commission Expires
NOTARY PUBLIC
With Office
STATE OF ALASKA

STATE OF ALASKA
SCIENTIFIC CRIME DETECTION LABORATORY

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N **130195**

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130195

APRIL 09, 2012

BMB 4/12/12

OPERATOR'S NAME:
ELY, RJ
OPERATOR'S NUMBER: 2681
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
VER OF CAL
O.L. #: 123456789
DEPT/AGENCY: CRG1
CASE/REPORT: 12-04XXXX
TEST TYPE: V
ALCO TARGET VALUE: .079
ALCO S/N: 75509

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.16 in

ALCO TARGET	.079	14:54
BLANK TEST	.000	14:55
INTERNAL STANDARD	VERIFIED	14:55
ALCO TV 30.16 in	.082	14:56
BLANK TEST	.000	14:56
SUBJECT SAMPLE	.000	14:57
BLANK TEST	.000	14:58
ALCO TV 30.16 in	.081	14:58
BLANK TEST	.000	14:59

BMB 4/12/12

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130195

APRIL 09, 2012
TIME 14:51

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	30.16 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~