



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130195

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130195

JULY 23, 2012

OPERATOR'S NAME:

ELY, RJ

OPERATOR'S NUMBER: 2681

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/NI :

VER OF CAL

O.L. #: 123456789

DEPT/AGENCY: CRG1

CASE/REPORT: 12-07XXXX

TEST TYPE: U

ALCO TARGET VALUE: .079

ALCO S/N: 75509

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.43 in  
ALCO TARGET .000 10:07  
BLANK TEST .000 10:08  
INTERNAL STANDARD VERIFIED 10:08  
ALCO TU 30.43 in .002 10:09  
BLANK TEST .000 10:09  
SUBJECT SAMPLE .000 10:10  
BLANK TEST .000 10:11  
ALCO TU 30.43 in .002 10:11  
BLANK TEST .000 10:12

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130195

JULY 23, 2012

TIME 10:05

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 30.43 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!""##%&'()\*+,-./0123456789:;<=>@ABCDEFGHIJK  
LMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
qrstuvwxyz{|}~

