

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

APR 30 2010

DataMaster cdm S/N <u>130195</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Robert J. Ely</u>	ID# <u>2681</u> Date <u>04-27-2010</u>
Agency <u>Craig Police Department</u>	Phone # <u>(907) 826-3330</u>
Instrument Location <u>506 2<sup>nd</sup> Street, Craig Alaska 99921</u>	
Alco S/N <u>X172978</u>	Target Value <u>.079</u> High Pressure <u>750</u> PSI
Alco Test Value Average <u>.081</u>	<u>.082</u>
1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature <u>Robert J. Ely</u>	

(OVER)

(Do Not write in the area below)

*fmb*  
5/3/10

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

*Nita J. Bolz*  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 8<sup>th</sup> day of June, 2010.

*Carolyn M. Noland*  
(Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130195

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130195

APRIL 27, 2010

OPERATOR'S NAME:  
ELY, RJ

OPERATOR'S NUMBER: 2681

SUBJECT'S LAST NAME:  
VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :  
VER OF CAL

O.L. #: 123456789

DEPT/AGENCY: CR61

CASE/REPORT: 2010-XXXXX

TEST TYPE: V

ALCO TARGET VALUE: .079

ALCO S/N: X172978

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.41 in

ALCO TARGET	.077	10:50
BLANK TEST	.000	10:59
INTERNAL STANDARD	VERIFIED	10:59
ALCO TV 29.41 in	.001	10:59
BLANK TEST	.000	11:00
SUBJECT SAMPLE	.000	11:01
BLANK TEST	.000	11:02
ALCO TV 29.41 in	.002	11:02
BLANK TEST	.000	11:03

*BMB 5/3/10*

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130195

APRIL 27, 2010  
TIME 10:56

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 43c

BAROMETER: 29.41 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>@ABCDEFGHIJKL  
MNPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz{|}~