

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130196</u> ✓
Name <u>JEAN ACHEE</u>		ID# <u>1470</u> Date <u>1/31/09</u> ✓
A	Agency <u>WASILLA POLICE DEPT.</u>	Phone # <u>352-5401</u>
Instrument Location <u>WASILLA POLICE DEPT.</u> ✓		
Alco S/N <u>X172942</u> ✓		Target Value <u>.078</u> ✓ High Pressure <u>1000</u>
B	Alco Test Value Average <u>.080</u> ✓	<u>.082</u> ✓
	1 st Alco	2 nd Alco
Signature <u>Jean Achee</u> ✓		<u>CS</u> <u>2/18/09</u>
(OVER)		


Under the Alaska Rules of Evidence, I certify that: (Do Not Write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Jean P Achee, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed.



Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

2/18/09

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130196 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130196

JANUARY 31, 2009 ✓ *CSB 2/18/09*

OPERATOR'S NAME: *80* PACHE, JEAN P.
OPERATOR'S NUMBER: 1470
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.I.L. #: 1234567890
DEPT/AGENCY: WACD
CASE REPORT: 09-3101
TEST TYPE: V
ALCO TARGET VALUE: .078 ✓
ALCO S/N: M172942

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.38 in		
ALCO TARGET	.078	00:02
BLANK TEST	.000	00:00
INTERNAL STANDARD	VERIFIED	00:04
ALCO TV 29.30 in	.000 ✓	00:05
BLANK TEST	.000	00:05
SUBJECT SAMPLE <i>CSB 2/18/09</i>	.000	00:05
BLANK TEST	.000	00:07
ALCO TV 29.38 in	.062 ✓	00:07
BLANK TEST	.000	00:08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130196

JANUARY 31, 2009 ✓
TIME 00:01

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 41c
BAROMETER: 29.38 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST
ALCO TV 29.38 in .062 ✓