

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130196

Name William Rapson ID# 6208 Date 4/3/08

A Agency WASILLA P.D Phone # 352-5417

Instrument Location WASILLA P.D

B Alco S/N X172966 Target Value .084 High Pressure 1000

Alco Test Value Average .086 .087
1st Alco 2nd Alco

Signature William Rapson CDB
4/8/08

(OVER)

Under the Alaska Rules of Evidence, I certify that: (Do Not write in the area below)

- (a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.
- (b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, William Rapson, whose name appears in my records, is qualified to verify the calibration of the above instrument.
- (c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz 4/12/08
Date
Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130196

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130196

APRIL 03, 2008

OPERATOR'S NAME:

RAPSON/WILLIAM/C

OPERATOR'S NUMBER: 6208

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 1234567

DEPT/AGENCY: WAS1

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .004

ALCO S/N: X172966

--- BREATH ANALYSIS ---

.004 ADJUSTED FOR 29.81 in
ALCO TARGET .003 15:05
BLANK TEST .000 15:06
INTERNAL STANDARD VERIFIED 15:06
ALCO TV 29.81 in .006 15:06
BLANK TEST .000 15:07
SUBJECT SAMPLE .000 15:08
BLANK TEST .000 15:08
ALCO TV 29.81 in .007 15:09
BLANK TEST .000 15:10

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130196

APRIL 03, 2008

TIME 15:11

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.81 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqr
stuvwxyz{|}~