

APR 17 2009

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130196</u> ✓	
Name <u>Brent Johnson</u>		ID# <u>4980</u>	Date <u>4/8/09</u> ✓
A	Agency <u>AST</u>	Phone # <u>883-5111</u>	
Instrument Location <u>Arctic Man - Summit Lake</u> ✓			
Alco S/N <u>X173004</u> ✓		Target Value <u>.081</u>	High Pressure <u>1100</u>
Alco Test Values		<u>.075</u> ✓ 1 st Alco	<u>.074</u> ✓ 2 nd Alco
Signature <u>[Signature]</u>			<u>[Signature]</u> 5/13/09

(OVER)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows. (Do not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

[Signature]
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 15 day of May, 2009.

[Signature] (Notary Seal Stamp)
 Carolyn M. Noland
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130196

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130196

APRIL 08, 2009

OPERATOR'S NAME:

JOHNSON BRENT W

OPERATOR'S NUMBER: 4988

SUBJECT'S LAST NAME:

VERIF

SUBJECT'S FIRST NAME/MI :

OF CAL

O.L. #: 123456

DEPT/AGENCY: TOKO

CASE/REPORT: 00-000

TEST TYPE: V

ALCO TARGET VALUE: .081 ✓

ALCO S/N: X173804 ✓

— BREATH ANALYSIS —

.081 ADJUSTED FOR 26.40 in
ALCO TARGET .071 20:17
BLANK TEST .000 20:18
INTERNAL STANDARD VERIFIED 20:18
ALCO TV 26.40 in .075 ✓ 20:18
BLANK TEST .000 20:19
SUBJECT SAMPLE .000 20:19
BLANK TEST .000 20:20
ALCO TV 26.40 in .074 ✓ 20:21
BLANK TEST .000 20:21

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130196

APRIL 08, 2009 ✓

TIME 20:13

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 26.40 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghi jklmno
pqrstuvwxyz{|}~"