

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

JUN 28 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130196</u>	
Name <u>Brent W. Johnson</u>		ID# <u>4980</u> ✓	Date <u>6/22/10</u> ✓
A	Agency <u>AWT - OPS</u>	Phone # <u>246-3307</u>	
Instrument Location <u>Plu Stinson</u>			
Alco S/N <u>2F</u> <sup>BMB</sup> <u>6/20/10</u>		Target Value <u>.080</u> ✓	High Pressure <u>1000</u>
B	Alco Test Values	<u>.084</u> 1 <sup>st</sup> Alco	<u>.084</u> 2 <sup>nd</sup> Alco
Signature <u>[Signature]</u>			

(OVER)

BMB  
7/12/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 26<sup>th</sup> day of July 2010.  
[Signature] (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130196

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130196

JUNE 22, 2010 *EMB 7/12/10*

OPERATOR'S NAME:  
JOHNSON BRENT W  
OPERATOR'S NUMBER: 4980  
SUBJECT'S LAST NAME:  
VERIF OF  
SUBJECT'S FIRST NAME/MI :  
CALIBRATION  
O.L. #: 000000  
DEPT/AGENCY: PVS1  
CASE/REPORT: 00-0000  
TEST TYPE: V  
ALCO TARGET VALUE: .000  
ALCO S/N: ~~130196~~ 2F *EMB 6/23/10*

*Spoke with Brent about Serial number.*

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.17 in		
ALCO TARGET	.000	16:16
BLANK TEST	.000	16:17
INTERNAL STANDARD	VERIFIED	16:17
ALCO TV 30.17 in	.004	16:17
BLANK TEST	.000	16:18
SUBJECT SAMPLE	.000	16:19
BLANK TEST	.000	16:19
ALCO TV 30.17 in	.004	16:19
BLANK TEST	.000	16:20

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130196

JUNE 22, 2010  
TIME 16:25

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	42c
BAROMETER:	30.17 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST  
!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJKL  
MNPQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
qrstuvwxyz{|}~?