

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130196</u>
Name	<u>William Rapson</u>	ID# <u>6208</u> Date <u>8/20/08</u>
A Agency	<u>WASILLA P.D.</u>	Phone # <u>352-5417</u>
Instrument Location		<u>WASILLA P.D.</u>
B Alco S/N	<u>X301568</u>	Target Value <u>.072</u> High Pressure <u>600</u>
Alco Test Value Average	<u>.077</u> <u>.070</u>	<u>.078</u>
	1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature	<u>William Rapson</u> ✓	<u>CVB</u> <u>8/22/08</u>
(OVER)		

Under the Alaska Rules of Evidence, I certify that: (Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, William Rapson, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

08/27/08

Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130196 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST      DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130196

AUGUST 29, 2008

OPERATOR'S NAME:  
RAPSON/WILLIAM/C  
OPERATOR'S NUMBER: 6200  
SUBJECT'S LAST NAME:  
TEST

SUBJECT'S FIRST NAME/MI :  
TEST

O.L. #: 1234567  
DEPT/AGENCY: WASH  
CASE/REPORT: TEST  
TEST TYPE: T - Voc  
ALCO TARGET VALUE: .072 ✓  
ALCO S/N: X301568 ✓

--- BREATH ANALYSIS ---

.072 ADJUSTED FOR 29.49 in  
ALCO TARGET .070 00:45  
BLANK TEST .000 00:47  
INTERNAL STANDARD VERIFIED 00:47  
ALCO TV 29.49 in .077 ✓ 00:47  
BLANK TEST .000 00:48  
SUBJECT SAMPLE .000 00:49  
BLANK TEST .000 00:49  
ALCO TV 29.49 in .078 ✓ 00:50  
BLANK TEST .000 00:51

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130196

AUGUST 29, 2008 ✓  
TIME: 08:52

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/28/01  
HEATERS  
SAMPLE CHANGER: 40c  
BREATH TUBE: 42c  
BAROMETER: 29.49 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF6  
HIJKLMNPOQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrstuvwxyz{|}~