

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130196 ✓

Supervisor/Operator Performing the Verification Procedure:
 Name Brandi Barnett ID# 7038 ✓ Date 9/8/10 ✓

A
 Agency SCDL Phone # 209-5695
 Instrument Location SCDL

B
 Alco S/N SCOTT09 ✓ Target Value 0.080 ✓ High Pressure 250
 Alco Test Values 0.084 ✓ 0.082 ✓
 1st Alco 2nd Alco


Signature Brandi M. Barnett NJB
12/14/10

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) **The referenced instrument is certified for evidentiary use in the State of Alaska.**


 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 16th day of DEC, 2010.

 (Notary Seal Stamp)

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130196

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130196

SEPTEMBER 08, 2010

OPERATOR'S NAME:

BARNETT/B

OPERATOR'S NUMBER: 7838

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: VOC

DEPT/AGENCY: AAPA

CASE/REPORT: VOC

TEST TYPE: Y

ALCO TARGET VALUE: 0.080

ALCO S/N: SCOTT08

— BREATH ANALYSIS —

.080 ADJUSTED FOR 29.93 in
ALCO TARGET .080 09:42
BLANK TEST .000 09:43
INTERNAL STANDARD VERIFIED 09:43
ALCO TV 29.93 in .004 09:44
BLANK TEST .000 09:44
SUBJECT SAMPLE .000 09:45
BLANK TEST .000 09:45
ALCO TV 29.93 in .082 09:46
BLANK TEST .000 09:46

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130196

SEPTEMBER 08, 2010

TIME 09:39

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.94 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstu
vwxyz{|}~