

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130196</u> ✓
Name <u>WILLIAM RAPSON</u>		ID# <u>6208</u> Date <u>9/19/08</u>
A	Agency <u>WASILLA P.D.</u>	Phone # <u>352-5417</u>
Instrument Location <u>WASILLA P.D.</u> ✓		
Alco S/N <u>70034</u> ✓ Target Value <u>.077</u> ✓ High Pressure <u>900</u>		
B	Alco Test Value Average <u>.084</u> ✓	<u>.085</u> ✓
	1 st Alco	2 nd Alco
Signature <u>William Rapson</u> ✓		<u>CS</u> <u>9/20/08</u>
(OVER)		

Under the Alaska Rules of Evidence, I certify that (write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, William Rapson, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

10/5/08

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130196 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130196

SEPTEMBER 19, 2008

OPERATOR'S NAME:

RAFSON/WILLIAM/C

OPERATOR'S NUMBER: 6208

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 123456789

DEPT/AGENCY: WASH

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .077 ✓

ALCO S/N: 70034 ✓

--- BREATH ANALYSIS ---

.077 ADJUSTED FOR 29.41 in		
ALCO TARGET	.075	07:56
BLANK TEST	.000	07:57
INTERNAL STANDARD	VERIFIED	07:57
ALCO TV 29.41 in	.004 ✓	07:57
BLANK TEST	.000	07:58
SUBJECT SAMPLE	.000	07:59
BLANK TEST	.000	07:59
ALCO TV 29.41 in	.005 ✓	08:00
BLANK TEST	.000	08:01

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130196

SEPTEMBER 19, 2008 ✓

TIME 07:54

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 40c

BREATH TUBE: 41c

BAROMETER: 29.41 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNPQRSTUVWXYZ[\]^_`abcde fghijklmnopq
rstuvwxyz{|}~