

# VERIFICATION OF CALIBRATION REPORT DEC 29 2009

*of DataMaster cdm Breath Test Instrument*  
 State of Alaska  
 Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130196</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Garrett Willis</u>	ID# <u>5560</u> Date <u>12/27/09</u> ✓
A Agency <u>AST</u>	Phone # <u>907-262-4453</u>
Instrument Location <u>Cooper Landing</u> ✓	
Alco S/N <u>57345</u> Target Value <u>.082</u> High Pressure <u>900</u>	
B Alco Test Values	
<u>.083</u> 1 <sup>st</sup> Alco	<u>.082</u> 2 <sup>nd</sup> Alco
Signature <u>Garrett Willis</u> ✓	
(OVER)	

*BMAB*  
1/5/10

- (Do Not write in the area below)
- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
  - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
  - (3) I am the Scientific Director of the State Breath Alcohol Program.
  - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
  - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
  - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

*Nita J. Bolz*  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 26<sup>th</sup> day of January, 2010  
*Carolyn M. Noland* (Notary Seal Stamp)

Carolyn M. Noland  
 Notary Public, State of Alaska  
 Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130196 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130196 ✓

DECEMBER 27, 2009

OPERATOR'S NAME:  
WILLIS/GARRETT/R  
OPERATOR'S NUMBER: 5560  
SUBJECT'S LAST NAME:

VOC  
SUBJECT'S FIRST NAME/NI :  
VOC

O.L. #: VOC  
DEPT/AGENCY: C000  
CASE/REPORT: VOC  
TEST TYPE: V  
ALCO TARGET VALUE: .082  
ALCO S/N: 57345

### — BREATH ANALYSIS —

.082 ADJUSTED FOR 29.60 in  
ALCO TARGET .081 07:17  
BLANK TEST .000 07:19  
INTERNAL STANDARD VERIFIED 07:19  
ALCO TV 29.60 in .083 ✓ 07:19  
BLANK TEST .000 07:20  
SUBJECT SAMPLE *gmb* .000 07:20  
BLANK TEST *15/110* .000 07:21  
ALCO TV 29.60 in .082 ✓ 07:21  
BLANK TEST .000 07:22

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130196 ✓

DECEMBER 27, 2009 ✓  
TIME 07:00

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: *GRW1* 01/08/09  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 29.60 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF6  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmno  
pqrstuvwxyz{|}~