

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130286 e

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130286

MAY 23, 2009 ✓

OPERATOR'S NAME: *COB 0126109*
NELSON, KEVIN HOON
OPERATOR'S NUMBER: 5349
SUBJECT'S LAST NAME:
UOC
SUBJECT'S FIRST NAME/MI.:
UOC
D.L. #: 0000000
DEPT/AGENCY: NPL1
CASE/REPORT: 09-0000
TEST TYPE: U
ALCO TARGET VALUE: .000 ✓
ALCO S/N: X172988 ✓

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.78 in ✓
ALCO TARGET .079 00:14
BLANK TEST .000 00:16
INTERNAL STANDARD VERIFIED 00:16
ALCO TV 29.78 in .001 ✓ 00:16
BLANK TEST .000 00:17
SUBJECT SAMPLE *COB 0126109* .000 00:18
BLANK TEST .000 ✓ 00:18
ALCO TV 29.78 in .000 ✓ 00:19
BLANK TEST .000 00:20

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130286

MAY 23, 2009 ✓
TIME 00:05

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/00/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 40c
BAROMETER: 29.83 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUN 07 2009

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N <u>130286</u> ✓	
Name <u>Sgt. Kevin Nelson</u>	ID# <u>5349</u> Date <u>5/23/09</u> ✓
A Agency <u>NORTH POLE PD</u> Phone # <u>(907) 408-0612</u>	
Instrument Location <u>125 Snowman Lane, North Pole, AK 99705</u>	
B Alco S/N <u>X172988</u> ✓ Target Value <u>.080</u> ✓ High Pressure <u>550 PSI</u>	
Alco Test Values	<u>.081</u> ✓ <u>.080</u> ✓
1 st Alco	2 nd Alco
Signature <u>[Signature]</u>	<u>[Signature]</u> 6/25/09

(OVER)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 26th day of June 2009.

[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

