

VERIFICATION OF CALIBRATION REPORT SEP 20 2010

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130286 ✓

Name Sgt. Kevin Nelson ID# 5349 ✓ Date 9/16/10 ✓

A Agency NORTH POLE PD Phone # (907)488-6902

Instrument Location NPPD, 125 SNOWMAN LN, NORTH POLE, AK 99705

B Alco S/N X124662 ✓ Target Value .087 ✓ High Pressure 800 PSI

Alco Test Values .085 ✓ .085 ✓

1st Alco 2nd Alco

Signature [Signature]

(OVER)

BMB
10/4/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 15th day of DEC, 2010.

[Signature] (Notary Seal Stamp)
 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130286

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

Breath Test Record			
Serial Number	130286	Model	DataMaster
Location	NPL1	North Pole Police Dept.	
Subject Information			
Last Name	Voc	O.L. #	0000000
First Name	Voc		
Operator Information			
Operator Name	Nelson, Kevin Hoon	Case	10-000
Operator #	5349	Department	NPL1
Test Results			
Alco Target	.087	Alco Serial	X124662
Nominal	0.086	Barometer	29.86
Dry Gas 1	0.085	Time	00:34:42
Subject Alcohol	0.000	Time	00:36:05
Dry Gas 2	0.085	Time	00:37:05
		Test Type	V
		Test Date	09/16/2010
		Test Time	00:33:23
		Report	Y

Diagnostic Test Record	
Serial Number	130286
Model	DataMaster
Location	NPL1
North Pole Police Dept.	
Diagnostic Test Results	
Test Date	09/16/2010
Test Time	00:31:07
Computer	Okay
Program	Okay
Software Date	01/08/2009
Sample Chamber (C)	49
Breath Tube (C)	42
Barometer (in)	29.86
Flow Detector	Okay
High Speed Pump	Okay
Detector	Okay
Filter	Okay
Quartz	Okay
Calibration	Okay

--- BREATH ANALYSIS ---

```

.087 ADJUSTED FOR 29.86 in
ALCO TARGET          .086      00:33
BLANK TEST           .000
INTERNAL STANDARD    VERIFIED  00:34
ALCO TU 29.86 in    .085      00:34
BLANK TEST           .000      00:35
SUBJECT SAMPLE       .000      00:36
BLANK TEST           .000      00:36
ALCO TU 29.86 in    .085      00:37
BLANK TEST           .000      00:37
    
```

```

PUMP HIGH SPEED : OKAY
FILTERS: OKAY
QUARTZ STANDARD
CALIBRATION
PRINTER TEST
    
```