

VERIFICATION OF CALIBRATION REPORT

NOV 08 2010

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DEC 29 2010

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130286 ✓

Name Off. Stan Swetzof ID# 5002 ✓ Date 11/1/10 ✓

A Agency North pole PD Phone # 907-488-0612

Instrument Location 125 Snowman Lane, North pole, AK 99705

B Alco S/N X124662 Target Value .087 High Pressure 700 PSI

Alco Test Values .082 .082
 1st Alco 2nd Alco

Signature [Signature] [Signature] CSB
10/30/10

(OVER)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 5th day of JAN, 2011, 2010.

[Signature] (Notary Seal Stamp)

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130286

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130286

NOVEMBER 01, 2010

OPERATOR'S NAME: *COB 12/30/10*
SNETZOF/STANLEY
OPERATOR'S NUMBER: 5002
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/NI :

L
O.L. #: L
DEPT/AGENCY: NP-1
CASE/REPORT:
TEST TYPE: U
ALCO TARGET VALUE: .067
ALCO S/N: X124662

--- BREATH ANALYSIS ---

.067 ADJUSTED FOR 29.73 in
ALCO TARGET .063 01:12
BLANK TEST .000 01:12
INTERNAL STANDARD VERIFIED 01:13
ALCO TO 29.73 in .062 01:13
BLANK TEST .000 01:14
SUBJECT SAMPLE *COB 12/30/10* .000 01:15
BLANK TEST .000 01:16
ALCO TO 29.73 in .062 01:16
BLANK TEST .000 01:17

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130286

NOVEMBER 01, 2010
TIME 01:09

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS
SAMPLE CHANGER: 43c
BREATH TUBE: 41c

BAROMETER: 29.73 in

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./:0123456789:;=<=>?@ABCDEFG
HIJL@PQRSTUWXYZ-!_ 'abcde fghijklmno
pqrstuvwxyz{|}~@