

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130286</u>
Name	<u>DC STAN SWETZOF</u>	ID# <u>5002</u> Date <u>4/19/11</u>
A Agency	<u>NORTH POLE PD</u>	Phone # <u>907-488-0612</u>
Instrument Location <u>125 Snowman Lane, North Pole, AK 99705</u>		
Alco S/N <u>X124662</u>		Target Value <u>.087</u> High Pressure <u>500 PSI</u>
Alco Test Values		
	<u>.085</u> 1 st Alco	<u>.085</u> 2 nd Alco
Signature <u>[Signature]</u>		
(OVER)		


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 15 day of JUNE, 2010. NOB

[Signature] (Notary Seal Stamp)
Daniel L. Nicholas Kyle Klopffer
Notary Public, State of Alaska
Commission Expires with Office



CRF
4/21/11

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130286

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130286

APRIL 19, 2011

CRF 4/21/11

OPERATOR'S NAME: SMETZOF/STANLEY
OPERATOR'S NUMBER: 5902
SUBJECT'S LAST NAME: VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI: L
O.I.L. #: L
DEPT/AGENCY: NPL1
CASE REPORT:
TEST TYPE: U
ALCO TARGET VALUE: .087
ALCO S/N: K124632

--- BREATH ANALYSIS ---

.087 ADJUSTED FOR 29.73 in		
ALCO TARGET	.086	04:53
BLANK TEST	.000	04:54
INTERNAL STANDARD	VERIFIED	04:54
ALCO TV 29.73 in	.085	04:54
BLANK TEST	.000	04:55
SUBJECT SAMPLE	.000	04:56
BLANK TEST	.000	04:56
ALCO TV 29.73 in	.085	04:57
BLANK TEST	.000	04:57

CRF 4/21/11

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130286

APRIL 19, 2011
TIME 04:49

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 43c
BREATH TUBE: 43c
BAROMETER: 29.73 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

0123456789:;@#%&'()*+,-./:;<=>?@BCDEFG
HIJKL/MNOPQRSTU VWXYZ[\]^_`abcde*ghijklmno
pqrstuvwxyzz{|}~