

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

JAN 13 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130 287
Name	Mitch LANGSETH	ID#	2612
		Date	1-6-10
A Agency	KENAI Police Dept	Phone #	283-2879
Instrument Location	Wildwood Pre Trial		
Alco S/N	70108	Target Value	.082
		High Pressure	1175
B Alco Test Value Average	.083		.083
	1 st Alco		2 nd Alco
Signature	Mitch Langseth		.083 1/13/10

(OVER)


(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 26th day of January 2010.


(Notary Seal Stamp)
Carol M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130287

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST



ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130287 ✓

JANUARY 06, 2010 ✓ *CPB 1/13/10*

OPERATOR'S NAME:
LANGSETH
OPERATOR'S NUMBER: 2612
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
Z
O.L. #: 0
DEPT/AGENCY: KEH1
CASE/REPORT: 0
TEST TYPE: U
ALCO TARGET VALUE: .082 ✓
ALCO S/N: 70108 ✓

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 29.78 in
ALCO TARGET .081 07:08
BLANK TEST .000 07:09
INTERNAL STANDARD VERIFIED 07:09
ALCO TU 29.78 in .083 ✓ 07:09
BLANK TEST .000 07:10
SUBJECT SAMPLE *CPB* .000 07:11
BLANK TEST .000 07:11
ALCO TU 29.78 in *1/13/10* .083 ✓ 07:12
BLANK TEST .000 07:13

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130287 ✓

JANUARY 06, 2010 ✓
TIME 07:06

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 42c
BAROMETER: 29.78 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~■