

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

MAR 11 2009

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130287</u>
Name	<u>Mitchell Langseth</u>	ID#	<u>2612</u> Date <u>3-9-09</u>
A Agency	<u>Kenai Police Dept</u>	Phone #	<u>283-7879</u>
Instrument Location	<u>Wildwood Patrol</u>		
B Alco S/N	<u>X301542</u>	Target Value	<u>.082</u> High Pressure <u>1200</u>
Alco Test Value Average	<u>.083</u>	<u>.083</u>	
	1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco	
Signature	<u>[Signature]</u>		

COB  
3/27/09

(OVER)

(Do Not write in the area below)  
Under the Alaska Rules of Evidence, I certify that,

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Mitch E Langseth, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

[Signature]

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

3/28/09

Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130287

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130287

MARCH 09, 2009 ✓ *COB 3/9/09*

OPERATOR'S NAME:  
LANGSETH

OPERATOR'S NUMBER: 2612

SUBJECT'S LAST NAME:  
UDC

SUBJECT'S FIRST NAME/MI:  
Z

OILC. #: 0

DEPT/AGENCY: KENI

CASE/REPORT: 0

TEST TYPE: U

ALCO TARGET VALUE: 002 ✓

ALCO S/N: X301542 ✓

--- BREATH ANALYSIS ---

.002 ADJUSTED FOR 30.24 in		
ALCO TARGET	.002	12:29
BLANK TEST	.000	12:30
INTERNAL STANDARD	VERIFIED	12:30
ALCO TV 30.24 in	.003 ✓	12:31
BLANK TEST	.000	12:32
SUBJECT SAMPLE <i>COB</i>	.000	12:32
BLANK TEST	.000	12:33
ALCO TV 30.24 in <i>COB</i>	.003 ✓	12:33
BLANK TEST	.000	12:34

*12:35*

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130287

MARCH 09, 2009 ✓

TIME 12:27

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/28/01

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 42c

BAROMETER: 30.24 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJKL  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrs tuvwxyz{|}~