

VERIFICATION OF CALIBRATION REPORT

MAY 05 2010

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: **DataMaster cdm S/N** 130287

Name Mitch LANGSETH ID# 2612 Date 4/29/10

A Agency KEVAI Police Dept Phone # 283-7879

Instrument Location w. Idluud Pretrial

B Alco S/N 70108 Target Value .082 High Pressure 500

Alco Test Value Average .084 .084

1st Alco 2nd Alco

Signature *Mitch Langseth* *CB*
5/24/10

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 29th day of June, 2010.

Carolyn M. Noland (Notary Seal Stamp)
 Carolyn M. Noland
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130 287

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130287

APRIL 29, 2010

OPERATOR'S NAME:

LANGSETH

OPERATOR'S NUMBER: 2612

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

Z

O.L. #: 0

DEPT/AGENCY: KENI

CASE/REPORT: 0

TEST TYPE: U

ALCO TARGET VALUE: .002

ALCO S/N: 70108

--- BREATH ANALYSIS ---

.002 ADJUSTED FOR 29.91 in
ALCO TARGET .001 15:10
BLANK TEST .000 15:11
INTERNAL STANDARD VERIFIED 15:11
ALCO TU 29.91 in .004 15:11
BLANK TEST .000 15:12
SUBJECT SAMPLE .000 15:12
BLANK TEST .000 15:13
ALCO TU 29.91 in .004 15:14
BLANK TEST .000 15:14

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130287

APRIL 29, 2010

TIME 15:15

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/31

HEATERS

SAMPLE CHAMBER: 40c

BREATH TUBE: 42c

BAROMETER: 29.91 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./:0123456789:;<=>?@ABCDEFGHIJKL
MNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~