

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N <u>130287</u> ✓	
Name <u>Mitch Langseth</u>	ID# <u>2612</u> Date <u>5/1/08</u>
A Agency <u>Kenai Police Dept</u>	Phone # <u>283-2829</u>
Instrument Location <u>Wildwood</u>	
B Alco S/N <u>X301572</u> ✓ Target Value <u>.083</u> ✓ High Pressure <u>1000</u>	
Alco Test Value Average <u>.084</u> ✓	
1 st Alco	2 nd Alco <u>.084</u> ✓
Signature <u>[Signature]</u>	<u>CSB</u> <u>5/5/08</u>

(OVER)

(Do Not write in the area below)
Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Mitch E Langseth, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed

[Signature] 5/5/08
Date
Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

MAY 02 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N ~~130287~~ 130287

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130287

MAY 01, 2008

OPERATOR'S NAME:

LANGSETH

OPERATOR'S NUMBER: 2612

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/MI :

VERIFICATION OF CAL

O.L. #: 0

DEPT/AGENCY: KEN1

CASE/REPORT: 0

TEST TYPE: V

ALCO TARGET VALUE: .083

ALCO S/N: K301572

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 29.91 in
ALCO TARGET .082 00:58
BLANK TEST .000 00:59
INTERNAL STANDARD VERIFIED 01:00
ALCO TV 29.91 in .084 01:00 ✓
BLANK TEST .000 01:01
SUBJECT SAMPLE .000 01:01
BLANK TEST .000 01:02 ✓
ALCO TV 29.91 in .084 01:02 ✓
BLANK TEST .000 01:03

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130287

MAY 01, 2008

TIME 00:55

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 42c
BAROMETER: 29.91 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!""#%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~■