

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

MAY 26 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130287			
Name	Mitchell Langseth	ID#	2612			
		Date	05/25/2010			
A	Agency	Kenai Police Department	Phone # (907) 283-7879			
	Instrument Location	Wildwood Pretrial Facility				
Alco S/N		75509	Target Value	.077	High Pressure	900
B		Alco Test Values	.077	.076		
		1 st Alco		2 nd Alco		
Signature					CGB 6/7/10	

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 24th day of June, 2010.

(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130287 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130287 ✓

MAY 25, 2010 ✓

OPERATOR'S NAME: LANGSETH
OPERATOR'S NUMBER: 2612
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
TEST

O.L. #: 0
DEPT/AGENCY: KEN1
CASE/REPORT: 0
TEST TYPE: U ✓
ALCO TARGET VALUE: .077 ✓
ALCO S/N: 75509 ✓

--- BREATH ANALYSIS ---

.077 ADJUSTED FOR 30.07 in ✓
ALCO TARGET .077 12:49
BLANK TEST .000 12:50
INTERNAL STANDARD VERIFIED 12:50
ALCO TU 30.07 in .077 12:51 ✓
BLANK TEST .000 12:52
SUBJECT SAMPLE .000 12:52 ✓
BLANK TEST .000 12:53 ✓
ALCO TU 30.09 in .076 12:53 ✓
BLANK TEST .000 12:54

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130287 ✓

MAY 25, 2010 ✓
TIME 12:44 ✓

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 42c
BAROMETER: 30.07 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abode/ghijklmno
pqrstuvwxyz{|}~!@